

CERTIFICATE OF COMPLIANCE

Certificate Number 20140430-E156931
Report Reference E156931-A3-UL
Issue Date 2014-APRIL-30

Issued to: MURATA POWER SOLUTIONS INC
11 CABOT BLVD, MANSFIELD MA 02048
UNITED STATES



This is to certify that representative samples of COMPONENT-Measuring, Testing and Signal Generation Equipment Components
See Addendum Page

Have been investigated by UL in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL 61010-1, (ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - Part 1: General Requirements), CAN/CSA-C22.2 No. 61010-1, (ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE - Part 1: General Requirements)

Additional Information: See the UL Online Certifications Directory at www.ul.com/database for additional information

Only those products bearing the UL Recognized Component Marks for the U.S. and Canada should be considered as being covered by UL's Recognition and Follow-Up Service and meeting the appropriate U.S. and Canadian requirements.

The UL Recognized Component Mark for the U.S. generally consists of the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory. As a supplementary means of identifying products that have been produced under UL's Component Recognition Program, UL's Recognized Component Mark: , may be used in conjunction with the required Recognized Marks. The Recognized Component Mark is required when specified in the UL Directory preceding the recognitions or under "Markings" for the individual recognitions. The UL Recognized Component Mark for Canada consists of the UL Recognized Mark for Canada:  and the manufacturer's identification and catalog number, model number or other product designation as specified under "Marking" for the particular Recognition as published in the appropriate UL Directory.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Recognized Component Mark on the product.



William R. Carney, Director, North American Certification Programs

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus



CERTIFICATE OF COMPLIANCE

Certificate Number 20140430-E156931
Report Reference E156931-A3-UL
Issue Date 2014-APRIL-30

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Power Meter-ACM20 Series Power Meter

AC Power Meter Series:

ACM20-X-AC1-Y-XX-C and ACM20-X-AC1-Y-Z-XX-C where "X" represents the Current Transformer Input with a value either 1 = 0-5 A, 2 = 0-10 A, 3 = 0-20 A, 4 = 100 A or 5 = 32 A; and where "Y" delineates the digit for the color of the LEDs such that R = red, B = blue, G = green; and where the letter "Z" is either "blank" for Power Factor or F for Frequency; and where "XX" (optional ;non safety related) are placeholders for 1=50.00 reading, 2=75.00 reading, 3=100.0 reading or 5=500.0 reading; and finally, C = RoHS.

ACM3P Series Ammeter

AC 3-phase Ammeter Series: ACM3P-X-AC1-Y-XX-C where "X" represents the Current Transformer Input with a value either 1 = 0-5 A, 2 = 0-10 A, 3 = 0-20 A, 4 = 100 A or 5 = 32 A; and where "Y" delineates the digit for the color of the LEDs such that R = red, B = blue, G = green; and where "XX"(optional ;non safety related) are placeholders for 1=50.00 reading, 2=75.00 reading, 3=100.0 reading or 5=500.0 reading; and finally, C = RoHS.



William R. Carney, Director, North American Certification Programs

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus

